



Case File Number: MH-23988-SU-XRE-001

Issuance Date: 11/06/2015

Document Number: 15-TA-46015

Expiry Date: 11/06/2016

NO OBJECTION CERTIFICATE FOR IMPORT OF MEDICAL DIAGNOSTIC X-RAY EQUIPMENT FOR OBTAINING TYPE APPROVAL

In exercise of powers conferred under Section 16 of the Atomic Energy Act, 1962 read in conjunction with Rule 3 of the Atomic Energy (Radiation Protection) Rules, 2004, the Atomic Energy Regulatory Board (AERB) hereby grants NOC for import, for obtaining type approval in favour of **MR. ANIL RAMACHANDRAN, CHAIRMAN AND MANAGING DIRECTOR, ANALYTICAL AND MEDICAL TECHNOLOGIES, NAVI MUMBAI** for the following X-ray equipment.

Type of equipment	Dental (Hand-held)
Model Name	DX3000
Manufacturer	DEXCOWINCO LTD, Korea
Maximum operating voltage	60.000 kVp
Maximum operating current	1.000 mA

Terms and Conditions

1. This equipment shall be supplied only to the institution which has obtained the requisite procurement permission from AERB.
2. For obtaining Type Approval, the performance of the equipment specified, shall be demonstrated to the representative(s) of AERB as per the technical documents submitted to AERB with respect to operational and radiation safety parameters.
3. Licensee shall be responsible for the installation, commissioning, servicing and maintenance of the equipment.
4. Licensee shall ensure that in case the performance of the above unit is not satisfactory for issuance of type approval certificate, the unit shall be returned to the Manufacturer with intimation to AERB.

Dr. Avinash U. Sonawane
Head

MR. ANIL RAMACHANDRAN
ANALYTICAL AND MEDICAL TECHNOLOGIES
102, SHRI AMBIKA HERITAGE, FIRST FLOOR, PLOT NO. 1, SECTOR 1, KHARGHAR NEXT TO KHARGHAR STATION NAVI MUMBAI, MAHARASHTRA-410210

